QUANSYS BIOSCIENCES
Quansys Biosciences, Inc. | 365 N 600 W Logan, UT 84321 | (435) 752-0531

CERTIFICATE OF ANALYSIS


Q-Plex™ Human Pneumococcal IgG (14-Plex)

REF	Part Number: 482249HU	Software Product Code: HPNM260113 23731 001
LOT	Lot Number: HPNK260113 24477 001	
	Expiration Date: 2027-MAY-08	

Reconstitute Calibrator using 500 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482252HU	HPNP260511	N/A	2027-MAY-11
Sample Diluent	481157HU	HPNS260928	N/A	2027-SEP-28
Calibrator	481155HU	HPNM260113	500 µL	2027-OCT-16
Detection	482156HU	HPND260508	N/A	2027-MAY-08
Substrate A	101159GR	QNSA260612	N/A	2028-APR-02
Substrate B+	101120GR	QNSB+260612	N/A	2028-APR-02
Wash Buffer	101158GR	QNSW260923	N/A	2028-SEP-23

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	1	0.80	µg/mL	0.00017	0.00034	0.21
2	3	0.13	µg/mL	0.00014	0.00028	0.12
3	4	0.32	µg/mL	0.000052	0.00010	0.11
4	5	0.70	µg/mL	0.00012	0.00025	0.31
5	6B	0.83	µg/mL	0.00014	0.00028	0.69
6	7F	0.78	µg/mL	0.00018	0.00037	0.21
7	8	1.19	µg/mL	0.00023	0.00045	0.62
8	9N	0.66	µg/mL	0.00011	0.00023	0.29
9	9V	0.58	µg/mL	0.00011	0.00021	0.15
10	12F	0.20	µg/mL	0.000034	0.000069	0.11
11	14	3.52	µg/mL	0.00075	0.0015	0.62
12	18C	0.64	µg/mL	0.00032	0.00064	0.14
13	19F	1.54	µg/mL	0.00058	0.0012	0.98
14	23F	0.51	µg/mL	0.00047	0.00094	0.35
15	CWPS	N/A	N/A	N/A	N/A	N/A
16	Control Spot	N/A	N/A	N/A	N/A	N/A
17	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: Kayla Kowallis Date: 190CT26

QA Release: K Cox Date: 190CT26

For Research Use Only