QUANSYS BIOSCIENCES

Quansys Biosciences, Inc. | 365 N 600 W Logan, UT 84321 | (435) 752-0531

CERTIFICATE OF ANALYSIS


Q-Plex™ Human Pneumococcal IgG (14-Plex) 30 mL Diluent

REF	Part Number: 482249HU-30	Software Product Code: HPNM260528 24138 001
LOT	Lot Number: HPNK260528 24138 001	
	Expiration Date: 2027-SEP-28	

Reconstitute Calibrator using 500 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482252HU	HPNP261012	N/A	2027-OCT-12
Sample Diluent	481157HU-30	HPNS260928	N/A	2027-SEP-28
Calibrator	482155HU	HPNM260528	500 µL	2027-OCT-21
Detection	482156HU	HPND261005	N/A	2027-OCT-05
Substrate A	101159GR	QNSA260917	N/A	2028-APR-02
Substrate B+	101120GR	QNSB+260917	N/A	2028-APR-02
Wash Buffer	101158GR	QNSW260728	N/A	2028-JUL-28

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	1	0.86	µg/mL	0.00013	0.00026	0.45
2	3	0.28	µg/mL	0.00013	0.00026	0.24
3	4	0.41	µg/mL	0.00015	0.00030	0.33
4	5	0.41	µg/mL	0.00017	0.00033	0.34
5	6B	0.50	µg/mL	0.00021	0.00041	0.29
6	7F	1.33	µg/mL	0.00022	0.00044	0.53
7	8	0.27	µg/mL	0.00011	0.00022	0.24
8	9N	0.72	µg/mL	0.00029	0.00057	0.38
9	9V	1.08	µg/mL	0.00017	0.00034	0.53
10	12F	0.37	µg/mL	0.00018	0.00035	0.29
11	14	0.52	µg/mL	0.000080	0.00016	0.23
12	18C	0.40	µg/mL	0.00013	0.00026	0.35
13	19F	3.67	µg/mL	0.0020	0.0040	2.51
14	23F	0.63	µg/mL	0.000093	0.00019	0.41
15	CWPS	N/A	N/A	N/A	N/A	N/A
16	Control Spot	N/A	N/A	N/A	N/A	N/A
17	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: Kaylee Kowallis Date: 21OCT26

QA Release: Courtney Noorda Date: 21OCT26

For Research Use Only