QUANSYS BIOSCIENCES
Quansys Biosciences, Inc. | 365 N 600 W Logan, UT 84321 | (435) 752-0531

CERTIFICATE OF ANALYSIS

Q-Plex™ Human Pneumococcal IgG (9-Plex) Aliquoted Bulk 30 mL Diluent


REF	Part Number: 482149HU-30-25
LOT	Lot Number: HPNK260528 24462 001
	Expiration Date: 2027-AUG-11

Software Product Code:
HPNM260528 24129 001

Reconstitute Calibrator using 500 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482152HU	HPNP260811	N/A	2027-AUG-11
Sample Diluent	481157HU-30	HPNS260928	N/A	2027-SEP-28
Calibrator	482155HU	HPNM260528	500 µL	2027-OCT-27
Detection	482156HU	HPND260811	N/A	2027-AUG-11
Substrate A	101159GR	QNSA260917	N/A	2028-APR-02
Substrate B+	101120GR	QNSB+260917	N/A	2028-APR-02
Wash Buffer	101158GR	QNSW261015	N/A	2028-OCT-15

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	CWPS	N/A	N/A	N/A	N/A	N/A
2	2	1.13	µg/mL	0.00027	0.00055	0.93
3	10A	1.34	µg/mL	0.00095	0.0019	0.95
4	11A	0.31	µg/mL	0.00017	0.00033	0.12
5	15B	1.73	µg/mL	0.00066	0.0013	0.49
6	17F	0.40	µg/mL	0.00018	0.00040	0.28
7	19A	0.85	µg/mL	0.00031	0.00061	0.35
8	20	1.17	µg/mL	0.00064	0.0013	0.33
9	22F	1.50	µg/mL	0.00073	0.0015	0.74
10	33F	1.35	µg/mL	0.00047	0.00094	0.29
11	Control Spot	N/A	N/A	N/A	N/A	N/A
12	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: Kayla Kowalski Date: 27OCT26

QA Release: K Cox Date: 27OCT26

For Research Use Only