QUANSYS BIOSCIENCES

Quansys Biosciences, Inc. | 365 N 600 W Logan, UT 84321 | (435) 752-0531

CERTIFICATE OF ANALYSIS


Q-Plex™ Mouse Cytokine Panel 1 HS (6-Plex)

REF	Part Number: 111349MS	Software Product Code: **MHSM230927-P16P**
LOT	Lot Number: MCP1K230927 24539 001	
	Expiration Date: 2027-MAY-17	

Reconstitute Calibrator using 550 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	111352MS	MCP1P260709	N/A	2027-JUL-09
Sample Diluent	112457MS	MHSS250917	N/A	2027-MAY-17
Calibrator	112455MS	MHSM230927	550 µL	2027-OCT-30
Detection	112456MS	MHSD260702	6 mL	2027-JUL-02
Streptavidin HRP	101173GR	QNSN260507	N/A	2027-NOV-07
Substrate A	101159GR	QNSA260612	N/A	2028-APR-02
Substrate B+	101120GR	QNSB+260612	N/A	2028-APR-02
Wash Buffer	101158GR	QNSW260707	N/A	2028-JUL-07

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	IFNγ	800	pg/mL	0.55	1.10	800
2	IL-1α	620	pg/mL	0.43	0.85	620
3	IL-1ß	2390	pg/mL	1.64	3.28	2390
4	IL-6	1110	pg/mL	0.76	1.52	1110
5	IL-10	520	pg/mL	0.36	0.71	520
6	TNFα	390	pg/mL	0.27	0.53	390
7	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: Kayla Kowalski Date: 30OCT26

QA Release: Courtney Noorda Date: 30OCT26

101896GR - 9.0

For Research Use Only